

New MRPCs constructions and tests
for 2017-2018 EEE Upgrade

New MRPCs test status

20170222001	leak problem < 1l/h	LAMP-01	not delivered
20170223002	leak problem (7 l/h)	LAMP-01	not delivered
20170225003	leak problem (2 l/h)	LAMP-01	not delivered
20170314004	OK	GENO-01	delivered
20170316005	OK	GENO-01	delivered
20170317006	OK	GENO-01	delivered
20170405007	OK	SIEN-02	delivered
20170406008	OK	SIEN-02	delivered
20170407009	OK	SIEN-02	delivered
20170425010	OK	CARI-01	delivered
20170426011	OK	CARI-01	delivered
20170427012	OK	CARI-01	delivered
20170509013	OK	TORI-05	delivered
20170510014	OK	TORI-05	delivered
20170511015	OK	TORI-05	delivered
20170523016	OK	LODI-03	delivered
20170524017	eff 86%	LODI-03	delivered
20170524018	eff 75%	LODI-03	delivered
20170719019	OK	spare - ROMA-01	
20170921020	OK	spare – FRAS-01	
20170926021	OK	CAGL-04	in transit
20170927022	OK	CAGL-04	in transit
20170928023	eff 87% - Re-Test	spare	
20171026024	OK	spare – COSE-01	delivered
20171121025	In test	BOLO-05	
20171123026	In test	BOLO-05	
20171124027	OK	BOLO-05	
20180221028	OK	CAGL-04	in transit
20180222029	eff 67% - Re-Test	spare	
20180227030	eff 71% - Re-Test	spare	
20180228031	OK	spare - SAVO-03	
20180320032	OK	spare	
20180322033	OK	spare	

Next Test

This week:

20171121025

20171123026

If Ok, these 2 are delivered, together with 27, to Bologna

Next week:

20180227030

20180222029

Find a possible reason of low efficiency -> Re-Test

20170928023